

EC CERTIFICATE

Full Quality Assurance System

Certificate No.:
10185-2017-CE-CZS-NA-PS Rev. 4.0

Project No.:
PRJC-89313-2008-PRC-CZE

Valid Until:
14 December, 2020

This is to certify that the quality system of:

MEDIN, a.s.

Vlachovická 619, 592 31, Nové Město na Moravě, Czech Republic

For design, production and final product inspection/testing of:

STERILE AND NON-STERILE MEDICAL DEVICES

Has been assessed with respect to:

THE CONFORMITY ASSESSMENT PROCEDURE DESCRIBED IN ANNEX II EXCLUDING SECTION 4 OF COUNCIL DIRECTIVE 93/42/EEC ON MEDICAL DEVICES, AS AMENDED

and found to comply.

Further details of the product(s) and conditions for certification are given overleaf.

Place and date:
Høvik, 27 July 2020



For:
DNV GL PRESAFE AS
Notified Body No.: 2460

Cathrine Wisbech

The certificate is digitally verified by blockchain technology. For more info, see www.dnvgl.com/assurance/certificates-in-the-blockchain.html



Certificate No.:
10185-2017-CE-CZS-NA-PS Rev. 4.0

Project No.:
PRJC-89313-2008-PRC-CZE

Valid Until:
14 December, 2020

Jurisdiction

Application of Council Directive 93/42/EEC of 14 June 1993, adopted as "Forskrift om Medisinsk Utstyr" by the Norwegian Ministry of Health and Care Services.

Certificate history:

Revision	Description	Issue Date
0.0	Supersedes DNV GL (NB0434) certificate No. 80567-2010-CE-CZS-NA 4.0 following transfer of notified body function to DNV Nemko Presafe AS (NB2460) Adding new product – Rib plates Removing products – Instruments for haemorrhoids and Cannulas, Dental endodontic instruments	01-02-2018
1.0	Removing products – Wizard Navigator and Rib stabilisation plate	08-03-2019
2.0	Recertification – Scope limited to "Rotary devices for surgery"	08-07-2020
3.0	Error in revision 3.0(NGPS), certificate not formally issued.	27-07-2020
4.0	Extension of validity of certificate revision 1.0, which expired June 14, 2020, with 6 months from the expiry date.	27-07-2020

Products covered by this Certificate:

Product Description	Product Name	Class
Reusable Surgical Medical Devices	1. Rotary devices for surgery	IIa
Dental Medical Devices	1. Endodony instruments 2. Endodony sterile instruments 3. Dental burs 4. Diamond burs	
Non-Active Implantable Medical Devices	1. Intramedullary nails 2. Bone screws 3. Auxiliary implants 4. Interference screws 5. Wires 6. Anchors 7. Staples 8. Reticule and baskets	IIb
Non-Active Implantable Medical Devices - Plates	1. Plates 2. Rib plates	

Certificate No.:
10185-2017-CE-CZS-NA-PS Rev. 4.0


Project No.:
PRJC-89313-2008-PRC-CZE

Valid Until:
14 December, 2020

The complete list of devices is filed with the Notified Body

Sites covered by this certificate

Site Name	Address
MEDIN, a.s.	Vlachovická 619, 592 31 Nové Město na Moravě, Czech Republic



Certificate No.:
10185-2017-CE-CZS-NA-PS Rev. 4.0

Project No.:
PRJC-89313-2008-PRC-CZE

Valid Until:
14 December, 2020

Terms and conditions

The certificate is subject to the following terms and conditions:

- Any producer (see 2001/95/EC for a precise definition) is liable for damage caused by a defect in his product(s), in accordance with directive 85/374/EEC, as amended, concerning liability of defective products.
- The certificate is only valid for the products and/or manufacturing premises listed above.
- The Manufacturer shall fulfil the obligations arising out of the quality system as approved and uphold it so that it remains adequate and efficient.
- The Manufacturer shall inform Presafe of any intended updating of the quality system and Presafe will assess the changes and decide if the certificate remains valid.
- Periodical audits will be held, in order to verify that the Manufacturer maintains and applies the quality system. Presafe reserves the right, on a spot basis or based on suspicion, to pay unannounced visits.

The following may render this Certificate invalid:

- Changes in the quality system affecting production.
- Periodical audits not held within the allowed time window.

Conformity declaration and marking of product

When meeting with the terms and conditions above, the producer may draw up an EC declaration of conformity and legally affix the CE mark followed by the Notified Body identification number of Presafe.

End of Certificate